Azer Scientific

01/09/2019

Certificate of Analysis

Product Name: Blood Bank Saline – Standard Solution

Description: Isotonic Saline Solution, 0.85% w/v, Non-buffered

Product No.: ES124410L

Lot. No.: 15380

Date of Manufacture: 01/04/2019

Expiration Date: 07/02/2020

Product Specifications:

Ingredients: Sodium Chloride, USP Grade
Deionized Water, NCCLS Type I Reagent Grade

Packaging: Product undergoes a two-phase sub-micron filtration process as well as UV Sterilization before filling process begins.
Packaged in 10 Liter Cubetainers (Manufactured by Hedwin Corporation.)

Materials:

	Part No.	Lot No.	Exp. Date
Container:	CUA0677	181113-1200	N/A
Spigot:	COM4401-IW	282246	N/A
Chemical:	ES1244-1000	9A15379	2020-07-02

QC Testing:

LOT NUMBER: 9A15379			
ATTRIBUTE	**SPECIFICATION**	**RESULTS**	**PASS / FAIL**
Appearance	Clear, colorless, no observable particulate	Pass	**PASS**
pH	6.5 - 7.0	6.8	**PASS**
Assay	0.85% to 0.90%	0.86%	**PASS**
Isotonicity (Macro observation)	No hemolysis	Pass	**PASS**
Isotonicity (Micro observation)	No RBC aggregates	Pass	**PASS**
TAMC	< 10 CFU/ml	<1 CFU/ml	**PASS**
TYMC	< 10 CFU/ml	<1 CFU/ml	**PASS**

Country of Origin: United States of America

Latex Free: This product contains no latex

Created By:

Colette Batog MT(ASCP)
Technical Support

701 Hemlock Road, Morgantown, PA 19543 610-524-5810 610-901-3046